

# GeneXpert Testing Procedure

## Global Procedure

Cartridge contains all necessary reagents: Just add the sample treated with the Sample Reagent

Mix "Sample Reagent" with the Sample.  
Incubate for 15 minutes at room temperature.



1

Transfer sample into the open port of the Xpert® MTB/RIF cartridge



2

Insert the cartridge and start the test



3

Handling Time = 2 MINUTES