

**FIND Evaluation of Bionote, Inc.**  
**NowCheck COVID-19 Ag Test**  
**External Report**  
*Version 1.6, 02 June 2022*

**Copyright and use of the report**

Copyright in this report is the property of FIND (or controlled by FIND). You are free to share, copy and redistribute the material in any medium or format provided that:

- (i) attribution: you must give appropriate credit to FIND and indicate if changes were made, you may do so in any reasonable manner, but not in any way that suggests that FIND endorses you or your use;
- (ii) non-commercial: you may not use the report for commercial purposes; and
- (iii) no derivatives: if you remix, transform, or build upon the materials or report, you may not distribute the modified materials or report unless with express authorization from FIND.

Presentation of data on our website does not impact any data ownership rights and FIND is not responsible for any use by any third party of these data. Data sources are provided.

**Evaluation process – private sector engagement**

FIND, the global alliance for diagnostics, seeks to ensure equitable access to reliable diagnosis around the world. It works closely with the private and public sectors and receives funding from donors and some of its industry partners. It has internal fire walls, policies and processes to protect it against any undue influence in its work or the publication of its findings.

More information on our policy and guidelines for working with private sector partners can be found here: <https://www.finddx.org/policies/>

For the COVID-19 response, FIND has commissioned independent evaluations of in vitro diagnostics following an Expression of Interest (EOI) process available on FIND’s website by which all test submissions were scored according to their regulatory status and time to market; the manufacturing and distribution capacity of the supplier; and the supplier-reported clinical and analytical performance.

**Document history**

Document Version	Date	Comment
1.0	18 September 2020	Initial release
1.1	16 October 2020	Corrected PCR sample type
1.2	3 November 2020	Updated Ct value sub analysis
1.3	5 November 2020	Analysis updated with inclusion of missing data

1.4	10 December 2020	Updated LOD methodology and results
1.5	20 April 2021	Updated Ease-of-use results
1.6	02 June 2022	Variants results added

## 1 Product info:

Manufacturer name	Bionote, Inc.
Test name	NowCheck COVID-19 Ag Test
Product code(s)	RG1901DG
Pack size(s)	25 Tests/Kit
Contents of kit	Test device, Extraction buffer tube, Nozzle cap, Swab, Paper stand, Film, Instructions for use
Equipment and consumables required, but not provided	Equipment: Timer Consumables: PPE
Product storage (temperature range)	2-30°C
Shelf-life (months)	24 months
Manufacturing site (country)	Republic of Korea

## 2 Study details:

Study design:	Prospective diagnostic evaluation studies across multiple, independent sites to determine the accuracy of COVID-19 antigen RDTs, using consecutive enrolment. Interim analyses are performed at 25% and 50% enrolment, and the evaluation is stopped if tests do not meet 97% specificity. Presence of symptoms, date of symptom onset and hospitalization status is collected for all enrolled participants.
Index assays:	Novel lateral flow format tests that detect recombinant SARS-CoV-2 antigens.
Reference method:	Results of the index test are compared to the routine, diagnostic RT-PCR result, which is used for clinical management
Limit of detection:	Analytical sensitivity, i.e. Limit of detection, was performed at the Liverpool School of Tropical Medicine in which standardized serial dilutions of cultured viral isolate were prepared. Proprietary swab provided in the kit was soaked in viral dilution series. Dilutions were tested in triplicate and the LOD was defined as the last dilution where all repeats were interpreted as positive.

Clinical performance:	<p>Sensitivity was calculated as the proportion of true positive results detected by NowCheck COVID-19 Ag Test among all positives by the reference method and reported as a percentage.</p> <p>Specificity was calculated as the proportion of true negative specimens, identified as negative by NowCheck COVID-19 Ag Test among all negatives by the reference method, and reported as a percentage.</p> <p>The 95% confidence intervals were calculated to assess the level of uncertainty introduced by sample size, using the Wilson's score method.</p>
Ease of use	<p>A System usability survey and ease of use questionnaire assessing the quality of the test, test preparation, ease of test execution, procedure time, ease of result interpretation, storage conditions and perceived settings of use was completed by operators and a final score out of 100 was calculated.</p>

### 3 Evaluation details:

Country of collaborator	Brazil
Location of clinical site(s) (city, town)	Marica, state of Rio de Janeiro
Health care level of site(s)	Community Testing Clinic
Study period (date to date)	30 July – 21 Aug 2020
Study cohort inclusion/exclusion	<p>Adults in community meeting national suspect definition</p> <p>Provided informed consent</p>
Sample type, antigen test	Nasopharyngeal swabs
Reference PCR method	<p>Lab-developed assay based on the US CDC protocol, which targets two regions (N1 and N2) of the nucleocapsid (N) gene of SARS-CoV-2 (<a href="https://www.fda.gov/media/134922/download">https://www.fda.gov/media/134922/download</a>)</p>
Sample type, PCR test	Nasopharyngeal swabs

### 4 Results:

#### 4.1 Study cohort

Country	<b>Brazil</b>
Total N	400

Age [mean (min-max), N]	40 (4-84), 396
Gender [%F, (n/N)]	54.7% (219/400)
Symptoms present [%Yes, (n/N)]	100%, (392/392)
Hospitalized (n, % Yes)	Not available
Days from symptom onset [median (Q1-Q3); N]	4 (3-6); 400
Days < 0-3 (n, %)	157, 39.3%
Days 4-7 (n, %)	185, 46.3%
Days 8+ (n, %)	58, 14.5%
Positivity [%, (n/N)]	25.5% (102/400)
PCR Ct [median (Q1-Q3); N]	23.1 (20.7-27.5); 102
Ct > 33 (n, %)	9 (8.8%)
Ct > 30 (n, %)	14 (13.7%)
Ct > 25 (n, %)	44 (43.1%)

#### 4.2 Estimation of clinical performance

Estimation of Clinical Performance Country	<b>Brazil</b>
Clinical Sensitivity (95% CI), N	89.2% (81.7, 93.9), 102
Sensitivity days ≤7, N	92.2% (84.8, 96.2), 90
Sensitivity Ct ≤ 33, N	91.4% (83.9, 95.6), 93
Sensitivity Ct ≤ 25, N	94.8% (85.9, 98.2), 58
Clinical Specificity (95% CI), N	97.3% (94.8, 98.6), 298
Invalid rate (% , n/N)	0%

\* LoD as reported in IFU:  $1.25 \times 10^{3.2}$  TCID<sub>50</sub>/ml, which is the equivalent of about  $1.4 \times 10^3$  pfu/ml

#### 4.3 Estimation of analytical performance

- Supplier reported LOD =  $3.12 \times 10^{2.2}$  TCID<sub>50</sub>/ml ~  $3.46 \times 10^2$  pfu/ml (Isolate BetaCoV/korea/KCDC03/2020, NCCP no. 43326)
- Verified LOD

Variant (lineage)	Lowest dilution detected	Verified LOD concentration	Viral Copy equivalent
-------------------	--------------------------	----------------------------	-----------------------

<b>UK wild type (B1)</b>	<b>2.5 x10<sup>2</sup> pfu/ml</b> ~ 3.525 x 10 <sup>2</sup> TCID <sub>50</sub> /ml	<b>2.5 x10<sup>2</sup> pfu/ml</b>	<b>5.9x10<sup>5</sup> genome copies/ml</b> applied to test
<b>Alpha (B.1.1.7)</b>	<b>1.0 x10<sup>2</sup> pfu/ml</b> ~ 1.41 x 10 <sup>2</sup> TCID <sub>50</sub> /ml	<b>1.0 x10<sup>2</sup> pfu/ml</b>	<b>6.3 x10<sup>3</sup> genome copies/ml</b> applied to test
<b>Gamma (P1)</b>	<b>1.0 x10<sup>2</sup> pfu/ml</b> ~ 1.41 x 10 <sup>2</sup> TCID <sub>50</sub> /ml	<b>1.0 x10<sup>2</sup> pfu/ml</b>	<b>4.3x10<sup>4</sup> genome copies/ml</b> applied to test
<b>Delta (B.1617.2)</b>	<b>5.0 x10<sup>1</sup> pfu/ml</b> ~ 7.05 x 10 <sup>1</sup> TCID <sub>50</sub> /ml	<b>5.0 x10<sup>1</sup> pfu/ml</b>	<b>8.1 x10<sup>4</sup> genome copies/ml</b> applied to test
<b>Omicron (BA.1)</b>	<b>5.0 x10<sup>2</sup> pfu/ml</b> ~ 7.05 x 10 <sup>2</sup> TCID <sub>50</sub> /ml	<b>5.0x10<sup>2</sup> pfu/ml</b>	<b>8.8x10<sup>4</sup> genome copies/ml</b> applied to test

*Note: viral dilution was applied directly to the test cassette, not to the provided swab*

#### 4.4 Ease of use

NowCheck COVID-19 Ag Test	<b>88</b> out of 100	3 operators, 1 country
---------------------------	----------------------	------------------------