

FIND Evaluation of SD Biosensor, Inc.

STANDARD F COVID-19 Ag FIA

External Report

Version 2.2, 30 September 2022

Copyright and use of the report

Copyright in this report is the property of FIND (or controlled by FIND). You are free to share, copy and redistribute the material in any medium or format provided that:

- (i) attribution: you must give appropriate credit to FIND and indicate if changes were made, you may do so in any reasonable manner, but not in any way that suggests that FIND endorses you or your use;
- (ii) non-commercial: you may not use the report for commercial purposes; and
- (iii) no derivatives: if you remix, transform, or build upon the materials or report, you may not distribute the modified materials or report unless with express authorization from FIND.

Presentation of data on our website does not impact any data ownership rights and FIND is not responsible for any use by any third party of these data. Data sources are provided.

Evaluation process – private sector engagement

FIND, the global alliance for diagnostics, seeks to ensure equitable access to reliable diagnosis around the world. It works closely with the private and public sectors and receives funding from donors and some of its industry partners. It has internal fire walls, policies and processes to protect it against any undue influence in its work or the publication of its findings.

More information on our policy and guidelines for working with private sector partners can be found here: <https://www.finddx.org/policies/>

For the COVID-19 response, FIND has commissioned independent evaluations of in vitro diagnostics following an Expression of Interest (EOI) process available on FIND's website by which all test submissions were scored according to their regulatory status and time to market; the manufacturing and distribution capacity of the supplier; and the supplier-reported clinical and analytical performance.

Document history

| Document version | Date | Comment |
|------------------|------------------|--------------------------------------|
| 1.0 | 16 October 2020 | Interim version |
| 2.0 | 20 October 2020 | Data for Germany added |
| 2.1 | 10 December 2020 | LoD Methodology updated; EoU updated |

| | | |
|-----|-------------------|---------------------------------|
| 2.2 | 30 September 2022 | LOD data for new variants added |
|-----|-------------------|---------------------------------|

1 Product info:

| | |
|--|--|
| Manufacturer name | SD Biosensor, Inc. |
| Test name | STANDARD F COVID-19 Ag FIA |
| Product code(s) | F-NCOV-01G, 10COV30D |
| Pack size(s) | 25 |
| Contents of kit | Test device (individually in a foil pouch with desiccant), Extraction buffer tube, Nozzle cap, Sterile Swab, Paper stand, IFU |
| Equipment and consumables required, but not provided | Equipment: STANDARD F Analyzer, Timer, refrigerator (optional, for storage of specimens prior to testing, if applicable) Consumables: PPE |
| Product storage (temperature range) | 2-30°C |
| Shelf-life (months) | <i>To be confirmed</i> |
| Manufacturing site (country) | Republic of Korea |

2 Study details:

| | |
|---------------------|---|
| Study design: | Prospective diagnostic evaluation studies across multiple, independent sites to determine the accuracy of COVID-19 antigen RDTs, using consecutive enrolment. Interim analyses are performed at 25% and 50% enrolment, and the evaluation is stopped if tests do not meet 95% specificity. Presence of symptoms, date of symptom onset and hospitalization status is collected for all enrolled participants. |
| Index assays: | Novel lateral flow format tests that detect recombinant SARS-CoV-2 antigens. |
| Reference method: | Results of the index test are compared to the routine, diagnostic RT-PCR result, which is used for clinical management |
| Limit of detection: | Analytical sensitivity, i.e. Limit of detection, was performed at the Liverpool School of Tropical Medicine in which standardized serial dilutions of cultured viral isolate were prepared. Proprietary swab provided in the kit was soaked in viral dilution series. Dilutions were |

| | |
|-----------------------|--|
| | tested in triplicate and the LOD was defined as the last dilution where all repeats were interpreted as positive. |
| Clinical performance: | <p>Sensitivity was calculated as the proportion of true positive results detected by STANDARD F COVID-19 Ag FIA among all positives by the reference method and reported as a percentage.</p> <p>Specificity was calculated as the proportion of true negative specimens, identified as negative by STANDARD F COVID-19 Ag FIA among all negatives by the reference method, and reported as a percentage.</p> <p>The 95% confidence intervals were calculated to assess the level of uncertainty introduced by sample size, using the Wilson's score method.</p> |
| Ease of use | A System usability survey and ease of use questionnaire assessing the quality of the test, test preparation, ease of test execution, procedure time, ease of result interpretation, storage conditions and perceived settings of use was completed by operators and a final score out of 100 was calculated. |

3 Evaluation details

| | | |
|---|---|---|
| Country of Collaborator | Brazil | Germany |
| Location of clinical site(s) (city, town) | <ol style="list-style-type: none"> 1. Macae, state of Rio de Janeiro 2. UFRJ | <ol style="list-style-type: none"> 1. Heidelberg 2. Berlin |
| Health care level of site(s) | <ol style="list-style-type: none"> 1. Community testing clinic 2. Tertiary level hospital | <ol style="list-style-type: none"> 1. Heidelberg: Drive-in testing Center 2. Berlin: Ambulatory testing clinic of Charité – University Hospital |
| Study period (date to date) | <ol style="list-style-type: none"> 1. Macae: 17 Aug – 9 Sept 2020 2. UFRJ: 11 Jul – 8 Aug 2020 | <ol style="list-style-type: none"> 1. Heidelberg: 15 June-18 July 2020 2. Berlin: 6 July – 23 Sept 2020 |
| Study cohort inclusion/exclusion | <p>Adults in community meeting national suspect definition</p> <p>Provided informed consent</p> | <p>Adults able to ambulate and meeting suspect definition of the Department of public health</p> <p>Provided informed consent</p> |
| Sample type, antigen test | Nasopharyngeal swab | <ol style="list-style-type: none"> 1. HD: Nasopharyngeal swabs 2. Berlin: Combined naso-/oropharyngeal swab |
| Reference PCR method | <ul style="list-style-type: none"> • Lab-developed assay based on the US CDC protocol, which targets two regions | <ul style="list-style-type: none"> • Cobas SARS-CoV-2 (Roche Diagnostics Inc) <ul style="list-style-type: none"> ○ N = 342 |

| | | |
|-----------------------|---|---|
| | <p>(N1 and N2) of the nucleocapsid (N) gene of SARS-CoV-2 https://www.fda.gov/media/134922/download);</p> <ul style="list-style-type: none"> n = 407 <ul style="list-style-type: none"> Lab-developed assay based on the Charité Universitätsmedizin Berlin protocol, which has 2 gene targets (E and RdRp) of SARS-CoV-2 https://www.who.int/docs/default-source/coronaviruse/protocol-v2-1.pdf); n = 46 | <ul style="list-style-type: none"> Abbott <i>RealTime</i> SARS-CoV-2 (Abbott Molecular, Inc) <ul style="list-style-type: none"> N = 1 Allplex 2019-nCov Assay (Seegene Inc) <ul style="list-style-type: none"> N = 20 LightMix® Modular SARS-CoV (COVID19) E-gene (Tib Molbiol) <ul style="list-style-type: none"> N = 233 Cobas (Roche) or Thermofisher (Multiplex TaqPath COVID-19 CE-IVD RT-PCR Kit) <ul style="list-style-type: none"> N = 80 |
| Sample type, PCR test | Nasopharyngeal swab | Nasopharyngeal (n=305), Naso/oropharyngeal (n=342) and/or Oropharyngeal swabs (n=32) |

4 Results

4.1 Study cohort

| Country | Brazil | Germany |
|---|-----------------------|------------------------|
| Total N (valid PCR results) | 453 | 676 |
| Age [mean (min-max), N] | 39 (0-95), 451 | 37.8 (18-85), 676 |
| Gender [%F, (n/N)] | 59% (268/453) | 54% (360/667) |
| Symptoms present [%Yes, (n/N)] | 93.6% (421/450) | 77.3% (517/669) |
| Hospitalized (n, % Yes) | Not available | Not available |
| Days from symptom onset [median (Q1-Q3); N] | 4 (3-6); 421 | 3 (2-5); 505 |
| Days < 0-3 (n, %) | 131, 31% | 257, 50.9% |
| Days 4-7 (n, %) | 248, 59% | 202, 40% |
| Days 8+ (n, %) | 42, 10% | 46, 9.1% |
| Positivity [%, (n/N)] | 26%, (120/453) | 5.77%, (39/676) |
| PCR Ct [median (Q1-Q3); N] | 23.4 (19.9-27.7); 120 | 24.7 (21.73-27.81); 39 |
| Ct > 33 (n, %) | 10, 8.3% | 3, 7.7% |
| Ct > 30 (n, %) | 20, 16.7% | 7, 18% |

| | | |
|----------------|-----------|-----------|
| Ct > 25 (n, %) | 54, 43.3% | 19, 48.7% |
|----------------|-----------|-----------|

4.2 Estimation of clinical performance

| Country | Brazil | Germany |
|----------------------------------|-------------------------|------------------------------------|
| Clinical Sensitivity (95% CI), N | 77.5% (69.2, 84.1), 120 | 69.2 (53.6, 81.4), 39 [§] |
| Sensitivity days ≤7, N | 80.2% (71.1, 86.7), 100 | 81.2 (64.7, 91.1), 32 |
| Sensitivity Ct ≤ 33, N | 80.9% (72.6, 87.2), 110 | 75 (58.9, 86.2), 36 |
| Sensitivity Ct ≤ 25, N | 87.9% (77.9, 93.7), 66 | 100 (83.9, 100), 20 |
| Clinical Specificity (95% CI), N | 97.9 (95.7, 99), 333 | 96.9 (95.2, 98), 637 |
| Invalid rate (%), n/N | 0%, (0/453) | 0%, (0/676) |

[§] 37/39 positives were tested using Roche Cobas, 1/39 using Seegene Allplex and 1/39 using TibMolbiol.

4.3 Estimation of analytical performance

- Supplier-reported LOD = 3.28×10^2 TCID₅₀/ml ~ inactivated SARS-CoV-2 (2019-nCoV) NCCP 43326/2020/Korea strain
- Verified LOD

| Variant (lineage) | Lowest dilution detected | Verified LOD concentration | Viral Copy equivalent |
|-------------------|--|----------------------------|--|
| UK wild type (B1) | 1.0×10^3 pfu/ml ~ 1.41 x 10 ³ TCID ₅₀ /ml | 1.0×10^3 pfu/ml | 2.4×10^6 genome copies/ml applied to test |
| Alpha (B.1.1.7) | 1.0×10^2 pfu/ml ~ 1.41 x 10 ² TCID ₅₀ /ml | 1.0×10^2 pfu/ml | 6.3×10^3 genome copies/ml applied to test |
| Gamma (P1) | 5.0×10^1 pfu/ml ~ 7.05 x 10 ¹ TCID ₅₀ /ml | 5.0×10^1 pfu/ml | 2.1×10^4 genome copies/ml applied to test |
| Delta (B.1617.2) | 5.0×10^1 pfu/ml ~ 7.05 x 10 ¹ TCID ₅₀ /ml | 5.0×10^1 pfu/ml | 8.1×10^4 genome copies/ml applied to test |
| Omicron (BA.1) | 5.0×10^2 pfu/ml ~ 7.05 x 10 ² TCID ₅₀ /ml | 5.0×10^2 pfu/ml | 8.8×10^4 genome copies/ml applied to test |

Note: viral dilution was applied directly to the test cassette, not to the provided swab

4.4 Ease of use

| | | |
|------------|---------------|---|
| STANDARD F | 82 out of 100 | 7 operators, Germany 5 operators, Brazil |
|------------|---------------|---|